

SPECIFICATION FOR INFLATABLE SPLINTS

1. The Inflatable Splints should provide constant and even pressure to an injury.
2. The Inflatable Splints are to be made with clear plastic and are perfect for temporary use on fractured, broken, crushed or burned limbs.
3. It should conform to any patient's individual shape, easy and quick to use patient's position to be easily re-set while splinted and have a push-pull valve to control pressure and should be cleaned and should be disinfected.
4. Should be X-Ray permeable.
5. Sizes of the air splints for leg 28" and hand 24" .