



BINOCULAR 8x30 R PRISM, OPTICAL INSTRUMENT


DRG. NO

COMD(I) 789/10 SH-7

FIRST ANGLE PROJECTION



1:1



CHAMFERS 0.1-0.3
 MAX. ALLOWABLE PYRAMIDAL ERROR 7'
 50.6 GLASSWAY = 32.20 AIRWAY
 $n_d = 1.57125 \pm 0.001$
 $v_d = 55.70 \pm 0.8\%$
 MATL. :- N-BAK 4/569560 SCHOTT

DT. - 18-06-2010 SCALE - NTS

	MADE BY: P.C.SINGH	MATL.	TREAT.
	TR-	AS ABOVE	AS ABOVE
	CH-	SHEETS. 2	SHT. NO. 1
DA-01HS dt-02/01/15	APPROVED	DRAWING NO.	
AMENDMENTS	 JWM/D.O.	MCT-2971/1	
OFD/DO/002			